

Claro™ Fasta DNA Prep Kit

HQM10.11-48

Claro™ Fasta DNA Prep Kit 48 rxn

Instructions for use


1. General Information (Bulk Kit)

The Claro™ Fasta DNA Prep Kit is a ready to use kit comprising two reagents enabling pre-analytical processing of a variety of sample types including whole blood, saliva, bacteria or animal cells or tissues. In 3 minutes, with no prior sample preparation, the Claro™ Fasta DNA Prep Kit allows you to go from cells to downstream nucleic acid investigations without the need for isolation or purification.



2. Materials Provided

Material Provided	Quantity	Number of Preps
Tube 1: Lysis Buffer	1 Tube	48
Tube 2: Wash buffer	1 Tube	



3. Storage Conditions

Tubes are shipped and stored at room temperature. The bulk kit format has been designed for high-throughput genomic applications. We advise that the buffers are not exposed to the environment for extended periods and/or continual short exposures, if this occurs then unused buffer not used within 48 hours after opening should be disposed of. If required the buffers can be pre-aliquoted and then stored for later use.

4. Applications

The product can be used on virtually any DNA containing sample. The nucleic acid released has been successfully applied to molecular biology techniques including PCR, sequencing and cloning.

The Claro™ Fasta DNA Prep Kit accelerates pre-analytical processing in the following targets:

- Human Samples: including whole blood (fresh/frozen, with and without heparin or EDTA), plasma, urine and buccal swabs
- Microbiological samples: including bacteria (*E.coli*, *S.aureus*, *P.aeruginosa*, *K. pneumoniae*), viruses (HBV/HCV) and parasites (plasmodium)

Instructions for Use continued

5. Applications

The Claro™ Fasta DNA Prep Kit is particularly suited to the following areas:

- Laboratory-free situations including point of care and field-based diagnostics, near-patient testing, and integrated sample-processing workflows.
- Forensic Analysis
- Veterinary research
- Genomics and biomarker analysis

6. Typical Protocol - Rapid processing of whole blood

Ensure samples have thawed completely before starting this procedure.

- 6.1. Add 30µl of whole blood to 150µl of Reagent 1 (or scale up for larger sample volume).
- 6.2. Incubate for one minute at room temperature.
- 6.3. Take 10µl of the above mixture and combine with 40µl of Reagent 2 (or scale up for larger sample volume).
- 6.4. Add appropriate volume into PCR master mix (e.g. 5µl per 25µl reaction).

For applications where enhanced sensitivity is required, the following protocol modifications can be used: Samples can be mixed with Reagent 1 at a 1:4 or 1:3 ratio to reduce sample dilution (See Table 1). Samples that have been processed in step 6.1 can be added to Reagent 2 at 1:3, 1:2 or 1:1 ratio to reduce sample dilution (See Table 2).

Table 1: Processing samples in Reagent 1

Sample Volume (µl)	Reagent 1 Volume (µl)	Ratio
30	120	1:4
60	240	1:4
90	360	1:4
30	90	1:3
60	180	1:3
90	270	1:3

Table 2: Washing samples in Reagent 2

Extract from Tube 1 (µl)	Reagent 2 Volume (µl)	Ratio
10	30	1:3
10	20	1:2
10	10	1:1



This product is for laboratory research use only.
CAUTION: Not for diagnostic use.