

**(주) 바이오디**

(우.14322) 경기도 광명시 하안로 60 광명SK테크노파크 A동 1102호  
T\_ 02.6264.3399 F\_ 02.6264.3400 E\_order@bio-d.co.kr



#### Fixed volume Micropipette

New designing for smooth activation, excellent ergonomics, easy maintenance and fully autoclavable.

Precision calibration  
Recalibration mechanism



Calibration adjustment with simple tools !



#### Fixed volume Micropipette :

Cat.No. 121°C Auto clavable	Code.No. 121°C Auto clavable	Cat.No.	Code.No.	Volume	Accuracy $\leq$ %	CV $\pm$ $\mu$ l
8011122	TF05H	8011132	TF05	5 $\mu$ l	2.0%	0.1
8011123	TF10H	8011133	TF10	10 $\mu$ l	1.0%	0.1
8011124	TF20H	8011134	TF20	20 $\mu$ l	1.0%	0.2
8011125	TF25H	8011135	TF25	25 $\mu$ l	1.0%	0.25
8011126	TF50H	8011136	TF50	50 $\mu$ l	1.0%	0.5
8011127	TF100H	8011137	TF100	100 $\mu$ l	1.0%	1.0
8011128	TF200H	8011138	TF200	200 $\mu$ l	0.8%	1.6
8011130	TF500H	8011140	TF500	500 $\mu$ l	0.8%	4.0
8011131	TF1000H	8011141	TF1000	1000 $\mu$ l	0.8%	8.0