


EC20 Pocket Conductivity Tester

Instruction Manual



APER A INSTRUMENTS, LLC

www.aperainst.com


EC20 Pocket Conductivity Tester Instruction Manual

1. Battery Installation

Please install batteries according to the following steps. *Please note the correct direction of battery installation:

The Positive Side (“+”) OF EVERY SINGLE Battery MUST FACE UP.

(WRONG INSTALLATION OF BATTERIES WILL CAUSE DAMAGE TO THE TESTER AND POTENTIAL HAZARDS!)



2. Keypad Functions

■ Short press----- < 2 seconds

■ Long press----- > 2 seconds

	<ol style="list-style-type: none">1. Short press to turn on, long press to turn off;2. When turned off, long press to enter setup;3. In mode setting, short press to change parameter;
	<ol style="list-style-type: none">1. When turned on, long press to enter calibration mode.2. In calibration mode, short press to confirm calibration;3. In mode setting, short press to confirm parameter selection.




Diagram - 1

3. Complete Kit







Diagram - 2

4. Calibration

4.1 Rinse the probe in distilled water and dry it with a clean tissue. Short press  to turn on the tester.

4.2 Long press  to enter calibration mode; Short press  to exit.










Diagram - 3

4.3 Dip the probe into 1413 $\mu\text{S}/\text{cm}$ calibration solution. Stir gently, leave it to stand, LCD displays  (Diagram 3), short press  to complete 1-point calibration. The tester returns to measuring mode, and calibration icon  appears at the bottom left of LCD.

4.4 Rinse probe in distilled water and dry it. Follow the steps in 4.2-4.3 to complete 2nd point calibration in 12.88 mS/cm calibration solution, tester returns to measurement mode, calibration icons   display on bottom left of LCD.

5. Conductivity Measurement

5.1 Short press  to turn on tester. Rinse probe in distilled water and dry it with a clean tissue to remove excess water.

5.2 Stir the probe in the sample solution gently, leave it to stand. Get readings after the smile icon comes up and stays.

5.3 Rinse off the tester thoroughly in distilled water after each test.

6. Notes



6.1 If it's first time use or the tester hasn't been used for a long time, we recommend soaking the probe in 12.88 mS/cm solution for 15-30 minutes to restore its sensitivity and speed up the response time.

6.2 The tester adopts 1413 μ S and 12.88 mS standard calibration solutions. Users can use 1-2 point calibrations as needed. For most circumstances, calibrating in 1413 μ S to complete 1st point calibration will meet testing requirements.

6.3 Unit Information: 1 mS/cm = 1000 μ S/cm

Once the readings are greater than 2000 μ S/cm, the unit will automatically become mS/cm.

6.3 The tester has self-diagnosis functions:

Symbol	Self-Diagnosis information	How to fix
<i>Er 1</i>	Wrong calibration solution, which exceeds the recognizable range of the meter.	1. Check if calibration solution is correct 2. Check if probe is damaged.
<i>Er 2</i>	 is pushed before measurement is stable (😊 comes up and stays)	Wait for the smile icon to stay, and then short press 







6.4 The tester has already been calibrated after manufacture. Usually users can use the tester right away, or test it in the calibration solutions to test its accuracy. When error is large, calibrate it before using.

7. Parameter Setting

7.1 Parameter setting reference chart:

Symbol	Parameter Setting content	Code	Factory Default
P1	Select conductivity ranges	Aut- μ S-mS	Aut
P2	Restore to factory default	No – Yes	No

7.2 How to setup parameters:

When turned off, long press  to enter setup → short press  to switch P1-P2 → Short press , parameter flickering → short press  to choose, short press  to confirm parameter selection → Long press  to switch off.

7.3 Parameter setting notes

a) Select conductivity ranges (P1):

Aut-- All ranges: 0 – 199.9 μ S/cm, 20 – 1999 μ S/cm, 2.0 – 20.0 mS/cm;

Factory default is Aut.




b) Restore to factory default:

Select Yes to restore the calibration to the theoretical values and parameter setting to original values. When meter's calibration or measurement performs abnormally, this function can be adopted so the meter goes back to factory default setting and then users can conduct calibration or take measurements again.

8. Technical Specifications

Conductivity	Range	0 – 200.0 $\mu\text{S/cm}$, 0 – 2000 $\mu\text{S/cm}$, 0 – 20.00 mS/cm
	Resolution	0.1/1 $\mu\text{S/cm}$, 0.01 mS/cm
	Accuracy	$\pm 1\% \text{ F.S} \pm 1 \text{ digit}$
	Calibration points	1-2 points auto calibration
	Automatic Temp. Compensation	0-50°C
Temp.	Range	0-50°C
	Resolution	0.1°C
	Accuracy	$\pm 0.5^\circ\text{C}$

9. Other Functions & Parameters

Indication of calibration points		Auto Power-off	Power-off in 8 minutes if no operation
Indication of stable measurements		Waterproof level	IP67, floats on water
Self-Diagnosis information	Er1, Er2	Power Supply	AAA batteries*4
Low battery reminder	 Flashes to remind to replace batteries	Battery Life	Up to 1000 hours of continuous operation
Dimensions/Weight	Instrument: 40*31*178mm/107g; Carrying case: 190*165*140mm/438g		

9. Warranty

We warrant this instrument to be free from defects in material and workmanship and agrees to repair or replace free of charge, at option of APERA INSTRUMENTS, LLC, any malfunctioned or damaged product attributable to responsibility of APERA INSTRUMENTS, LLC for a period of **two years** from the delivery (a **six-month** limited warranty applies to electrodes). Warranty period is the time limit to provide free service for the products purchased by customers, not the service life of the tester or electrodes.

This limited warranty does not cover any damages due to:

- i. transportation;
- ii. storage;
- iii. improper use;
- iv. failure to follow the product instructions or to perform any preventive maintenance;
- v. modifications;
- vi. combination or use with any products, materials, processes, systems or other matter not provided or authorized in writing by us;
- vii. unauthorized repair;
- viii. normal wear and tear; or
- ix. external causes such as accidents, abuse, or other actions or events beyond our reasonable control.

(주) 바이오디

(우,14322경기도 광명시 하안로 60 광명테크노파크 A동 1102호

T_02.6264.3399 F_ 02.6264.3400 제품문의 : cs@bio-d.co.kr 업체문의 : ds@bio-d.co.kr



검색창 "바이오디" 검색

